



HGI PROFICIENCY TESTING

REPORT NINETY-THREE

Revision: 00

Final report

MARCH 2021

LABORATORY CODE: c

SCHEME CO-ORDINATOR: S MAMABOLO

CHECKED BY: R BABOOLAL (SCHEME MANAGER)


Disclaimer: Opinions and interpretations expressed herein are outside the scope of SANAS accreditation

THINKING QUALITY, QUALITY THINKING

REGISTRATION NUMBER: 2006/149731/23 (RMB INDUSTRIAL STATIONERS cc t/a)
81 CERAMIC CURVE, ALTON, RICHARDS BAY 3900 | TEL: +27(0)35 751 2446 | CEL: +27(0)83 6500151 | FAX: 0862605793
E-MAIL: ravi@coalconcepts.co.za

EXECUTIVE SUMMARY

1. Twenty samples were sent to participants with 19 result submissions.
2. Results indicated by * are out of the reproducibility limit of ± 5 units, but not flagged as outliers by Grubbs estimate.
3. Robust statistics were applied to calculate the robust standard deviation and robust average, the distribution of results did follow a Gaussian curve (below).



4. The trending of z-scores over time is a good indication of the laboratory's performance. Trending for your laboratory is as follows:

COMMENT BASED ON LAST FIVE Z-SCORE RESULTS

TABLE OF CONTENTS

ITEM	PAGE NUMBER
Letter to participant	4
List of participants	5
Type of sample used	6
Preparation of sample	6
HGI results and trend	7
Conclusion	8
Terms & Conditions	9

Dear Participant

RE: HGI PROFICIENCY TESTING RESULTS FOR THE MONTH OF MARCH

Thank you for your participation in the Coal Concepts HGI proficiency testing scheme.

Your laboratory code is as per the cover page.

All results are totally confidential. Any results in bold, italics and underlined are outliers. Where applicable, the most extreme outliers have been eliminated from calculations, using the Grubbs estimate for outliers. Please take note of the following: **(Please see conclusion before accepting z-score below)**

1. Z-scores between -1 and +1 is deemed acceptable
2. Z-scores between -2 and -3 should serve as a warning that the analysis result could get worse
3. Z-scores between +2 and +3 should also serve as a warning that analysis results could get worse.
4. Z- scores lower than -3 and exceeding +3 should warrant an investigation
6. All calculations can be made available upon request

The Coal Concepts scheme adheres to the requirements of ISO/IEC 17043:2010 – Conformity assessment – General requirements for proficiency testing.

Statistical analysis has been carried out using ISO/IEC 13528:2015-Statistical methods for use in proficiency testing by interlaboratory comparisons

Please find results attached together with Z-score trends.

Best Regards

R Baboolal

LIST OF PARTICIPANTS IN ALPHABETICAL ORDER

Alfred H Knight - Richards Bay Laboratory
Bureau Veritas Beira Laboratory
Bureau Veritas Nacala Laboratory
Bureau Veritas Inspectorate Laboratories Alton
Bureau Veritas Inspectorate Laboratories Middelburg
Eskom Holdings Arnot Power Station
Eskom Erid : Research & Development
Eskom Holdings Hendrina Power Station
Eskom Holdings Kriel Power Station
Eskom Holdings Kendal Power Station
Eskom Holdings Matla Power Station
Eskom Holdings Grootvlei Power Station
Eskom Holdings Lethabo Power Station
Eskom Holdings Majuba Power Station
Eskom Holdings Matimba Power Station
Eskom Holdings Tutuka Power Station
Exxaro Grooteguluk
Noko Analytical Services - Witbank Laboratory
Ronewa Lab
SABS Richards Bay
Sibonisiwe
Siza Coal Services Middelburg
Umzamo Analytical Services – Witbank Laboratory

1. TYPE OF SAMPLE USED


The coal used in this proficiency testing round was bituminous coal from the Mpumalanga region of South Africa

2. PREPARATION OF SAMPLE

Approximately 350kg's of sample with an approximate topsize of 50mm was sourced. This was crushed to -4.75 mm using a jaw crusher. The 4.75mm screen was placed on a 1.18mm screen and the -4.75 mm material screened in batches of about 5kgs. Coal passing through the 4.75mm screen but retained on the 1.18mm screen was placed in a mixing drum. Once all the coal was screened and transferred to the mixing drum, it was mixed for approximately 4 hours. The material was then transferred to containers capable of holding about 1kg of coal sample. 300 samples were obtained in this way.

3. HGI RESULTS AND TRENDING

COAL CONCEPTS - PROFICIENCY TESTING - MARCH 2021			
ANALYTICAL PARAMETER : HARDGROVE INDEX			
	LAB ID	HGI	Z-SCORE
	1c	50	-0,73
	3c	54	0,00
	4c	52	-0,30
	7c	54	0,00
	9c	50	-0,71
	10c	53	-0,12
	11c**	46	-1,51
	13c	59	1,05
	14c	57	0,70
	15c**	32	-4,33
	19c**	61	1,41
	21c	50	-0,71
	22c	51	-0,42
	24c**	47	-1,31
	25c	51	-0,42
	26c	54	0,00
	27c**	105	10,38
	29c**	48	-1,08
	30c	57	0,62
	38c	57	0,70
	39c**	66	2,5
Number of results	-	21	-
OUTLIERS	-	2	-
AVERAGE	-	54	-
STD DEVIATION	-	5	-
MEDIAN	-	53	-
MAXIMUM	-	66	-
MINIMUM	-	46	-
RANGE	-	20	-
ROBUST AVERAGE	-	53	-
ROBUST STD DEVIATION	-	5	-
UoM	-	1	-
REPRODUCIBILITY			
UPPER REPRODUCIBILITY LIMIT	-	59	-
LOWER REPRODUCIBILITY LIMIT	-	49	-



4. CONCLUSION

4.1 The range is 20 which is smaller than the previous round's results.

4.2 Two outliers were detected using Grubbs estimate

4.3 The standard deviation is 5 units which is within reproducibility of 5 units.

4.4 The median, average and robust average values are similar, indicating that the extreme results do not affect the centralised values.

4.5 The overall Z- Score trend is evenly distributed

4.6 Homogeneity

SAMPLE NO.	TEST PORTION 1	TEST PORTION 2	sample av (Xt)	range (Wt)	range sqd
1	16,48	16,80	16,64	0,32	0,1024
2	16,68	16,70	16,69	0,02	0,0004
3	16,68	16,80	16,74	0,12	0,0144
4	16,61	16,71	16,66	0,10	0,0100
5	16,78	16,80	16,79	0,02	0,0004
6	16,55	16,80	16,68	0,25	0,0625
7	16,53	16,77	16,65	0,24	0,0576
8	16,53	16,81	16,67	0,28	0,0784
9	16,68	16,64	16,66	0,04	0,0016
10	16,68	16,76	16,72	0,08	0,0064
GENERAL AVERAGE			16,69		
STANDARD DEVIATION			0,047		
WITHIN SAMPLE STANDARD DEVIATION			0,129		
BETWEEN SAMPLE STANDARD DEVIATION			0,078		

4.7 Stability

SAMPLE NO.	TEST PORTION 1	TEST PORTION 2	sample av (Xt)	range (Wt)	range sqd
11	16,27	16,92	16,60	0,65	0,4225
12	16,31	16,92	16,62	0,61	0,3721
13	16,66	16,74	16,70	0,08	0,0064
14	16,45	17,05	16,75	0,60	0,3600
15	16,77	16,90	16,84	0,13	0,0169
16	16,56	16,67	16,62	0,11	0,0121
17	16,52	17,03	16,78	0,51	0,2601
18	16,27	16,66	16,47	0,39	0,1521
19	16,44	17,03	16,74	0,59	0,3481
20	16,64	16,68	16,66	0,04	0,0016
GENERAL AVERAGE			16,67		
STANDARD DEVIATION			0,107		
WITHIN SAMPLE STANDARD DEVIATION			0,312		
BETWEEN SAMPLE STANDARD DEVIATION			0,193		

The HGI samples were confirmed to be sufficiently homogenous and stable.

End of Report

COAL CONCEPTS: Terms and ConditionsReturn of results:

Laboratories participate in proficiency testing programs on the understanding that they will be sharing their results and information **anonymously** with other laboratories performing the same analysis. No return of results compromises the spirit of the programs, and reports will not be sent to laboratories unless they return results. Payment in full is required from all laboratories enrolling whether they return results or not.

Errors in Participant Proficiency Testing Results:

Proficiency testing reports should reflect the level of accuracy that a regular testing client would receive.

If a participant finds an error in their proficiency testing results, they may notify us in writing and change their submission **PRIOR** to the due date for return. Changes after this time will not be accepted.

Coal Concepts' reports results *as submitted* by participants.

On occasion, it seems as though participants have mixed up the samples or not processed the samples according to the instructions. Coal Concepts cannot make assumptions of this nature and change results 'to suit'. We also cannot compromise the integrity of the programs by suggesting to some participants that they should review their results prior to the due date. (This is unfair to other participants) It is the responsibility of the participants to check all aspects of the program, including sample identification, preparation, testing instructions, calculations and reporting of the results prior to results submission.

If samples are not in good condition on arrival to the participant laboratory, Coal Concepts must be notified in writing IMMEDIATELY, as often samples can be replaced in good time. Claims about samples received in bad condition will not be accepted after the report has been issued.

Late Enrolments and Late Results:

Late enrolment requests cannot always be accommodated, as sample manufacture must be scheduled well in advance to the shipping date of the program to allow all necessary quality assurance activities to be carried out.

Shipping of PT materials and evaluating test results from PTPs out of cycle with the mainstream programs is considerably time consuming and therefore costly. In order not to disadvantage participants able to comply with time frames, Coal Concepts may charge a late fee in the following circumstances:

Requests that Coal concepts staff enters results on behalf of participants

Requests to record results after the due date

Requests for PTP participation that is out of cycle with the scheduled dates

Shipping fees and Customs clearance:

Costs incurred for shipping samples and clearance of same through customs are the responsibility of the participating laboratory unless otherwise indicated

Non-payment of fees:

Coal Concepts retains the right to withhold reports and/or test materials and services when invoices are outstanding.

Confidentiality of results:

All data and information received by Coal Concepts from its clients are considered confidential unless the client has given express permission to pass on information.

Definitions:

The dictionary definitions of "collusion" and "falsification" are as follows.

· *Collusion*: A secret agreement or cooperation for a fraudulent or Deceitful purpose.

· *Falsification*: Deliberately changing something to be false. In proficiency testing terms, collusion is comparing data (and perhaps changing data) to fit in with a believed "correct" result. This is contrary to the spirit of proficiency testing programs, which are issued with the intention of providing an objective comparison of a laboratory's performance with others. Coal Concepts tries to minimise the occurrence of collusion by being aware that laboratories should be objective when they report their results, and should therefore not know the intended results at the time they are reporting to us.

Answers are not provided to clients until results have been submitted.

To prevent collusion and falsification our advice to clients is:

DON'T confer with others about PT samples or results.

DO accept the fact that everyone makes errors.

DON'T average the results or opinions of every person in the laboratory before selecting the answer to be submitted. Instead, use one of the answers AS SUBMITTED to you and take advantage of the Coal Concepts internal QA services and submit all answers generated by the technicians.

DO have confidence in your own results.

Proficiency Testing (PT) is a compulsory part of laboratory accreditation, but it is also an important tool for giving you confidence in your results. "Enhancing" your PT results with assistance from another participant cannot increase

Confidence in your laboratory's performance.

Coal concepts' testing staff are not told what the expected results are, nor what we are expecting.

We subject ALL results to analysis, even if they are different.

The staff have the right to check that the results we enter on their behalf are correctly transcribed.

Clients are always welcome to contact Coal Concepts to seek advice or information about collusion or falsification of data.

Policy for Participant Appeal of PT Performance Assessment:

If participants disagree with their performance assessment in a proficiency report, they should inform Coal Concepts in writing.

The response will include Coal Concepts interpretation of the outcome of the reassessment and an explanation of that outcome. (For example, explanation of a calculation, or the rationale for the outcome of the evaluation.)

If a mistake has been made by Coal Concepts, it will be dealt with via Coal Concepts' non-conformance system.

Liability

In no event shall a party's liability to the other party for direct damages exceed an amount equal to the value of the amount for the PT Programme, under that specific month